

QUICK REFERENCE GUIDE



SINCE 1964


This procedure guides you through a typical data collection process. If you need additional detail, consult the Users Help Guide in the software.

The M.O.L.E.[®] Thermal Profiler depends on the MAP (Machine-Assembly-Process) software to control how it collects and interprets data. Several kinds of data runs may need to be performed to achieve desired information, or the same data run may be performed repeatedly over time to monitor one process. Either way, each data run must be set up at least once.

The MAP software includes wizards that help you get started quickly, even if you are a beginner or infrequent user. With settings for the five likely -workflows, you can get up and running in no time. This Guide documents the process used with the **Fresh Start** workflow wizard.

Step I: Set MAP information

1) Double-click the M.O.L.E.[®] MAP application icon to start the software.



2) Using the USB cable, connect the M.O.L.E.[®] Thermal Profiler to the computer.

3) Make sure the M.O.L.E.[®] Thermal Profiler Power Pack battery is fully charged. When a M.O.L.E.[®] Thermal Profiler is selected, the software status bar displays the current battery voltage.

4) On the **File** menu, click **New**. The Start dialog box appears with the five workflow wizard options.





When navigating through the wizard, the step list on the left of the dialog box uses a color key to inform the user of the progression through the wizard.



Current




Completed



Remaining

5) On the **Start** dialog box, click the **Fresh Start** command button and the workflow wizard appears.

6) Select the desired instrument from the list box to make active. If there are none listed, click the **Scan for Instruments** command button to detect all available instruments.



7) Click the **Next** command button.

8) Select a machine from the drop-down list. If the desired machine does not appear in the list, click the **New** command button to create a new machine.

9) Set the machine recipe settings, such as Conveyor Speed, Zone Temperatures and Temperature units.

The screenshot shows the 'Fresh Start' software window. On the left is a sidebar with a tree view containing the following items: Select Instrument, Set Machine Information, Set Assembly Information, Set Process Information, Set Recording Parameters, Verify Instrument Status, Perform Data Run, and Read Data Run. The main area is titled 'Machine Selection' and includes a dropdown menu for 'Machines' (currently showing 'TPS_MD-200'), 'Heating Zones: 9', 'Cooling Zones: 1', and 'Length Units: cm'. There are 'New...' and 'Edit...' buttons. Below this is the 'Recipe Settings' section, which includes 'Conveyor Speed: 216.00 in/min', a 'Name' field (showing 'TPS_MD-200_Recipe1'), and buttons for 'Load...', 'Save...', 'Print...', 'Send to machine', and 'Notes...'. Further down, there are checkboxes for 'Zones Same Speed' and 'Top and Bottom Setpoints are the same', and a 'Temp Units: C' dropdown. At the bottom is a table with 10 columns (numbered 1 to 10) and 3 rows (Top Temp, Length, and Convection). The 'Top Temp' row contains values: 260, 300, 300, 420, 490, 575, 635, 637, 928, 102. The 'Length' row contains values: 80.00, 80.00, 80.00, 40.00, 40.00, 40.00, 40.00, 30.00, 20.00, 300.00. The 'Convection' row is empty. At the very bottom are buttons for 'Help', '<< Previous', 'Next >>', 'Finish', and 'Cancel'.

	1	2	3	4	5	6	7	8	9	10
Top Temp	260	300	300	420	490	575	635	637	928	102
Length	80.00	80.00	80.00	40.00	40.00	40.00	40.00	30.00	20.00	300.00
Convection										



The software includes features to save and load machine recipe setting files (*.OVS). These files are helpful so the user can quickly recall machine information and ensure it is always the same.

10) Click the **Next** command button.

11) Enter the assembly information such as part number, test product size, sensor locations and a product image.

Fresh Start

Assembly Information

Part Number: Notes...

Image File: Browse...

(X) Length: (Y) Width: Thickness: Units:

Load... Save...

Channel	On	Location	X	Y	Z
1	☒	Sensor 1 Location	0.00	2.50	0.00
2	☒	Sensor 2 Location	0.00	5.00	0.00
3	☒	Sensor 3 Location	0.00	7.50	0.00

IMAGE NOT AVAILABLE

Assembly Flow Enlarge...

Help << Previous Next >> Finish Cancel

12) Click the **Next** command button.


13) Unwind the sensor leads and insert the connectors into the M.O.L.E.[®] Thermal Profiler. Make sure to connect them to the same channel(s) that were configured in the wizard workflow.

14) Locate the thermocouple sensors on a test solar cell.



When locating a thermocouple sensor on a test solar cell, make sure it is slightly curved creating the proper tension required to remain in place.

15) Select your process specification. The user can select a standard or user created **Paste** from the database. Then choose the **Profile Type** (Ramp-Soak-Spike or Ramp-to-Spike). If your Paste does not appear in the database list click the **New** command button to create a new one.



When the user selects a paste from the database, they can use the radio buttons below the drop down box to filter the user created pastes from the master paste specifications.

16) Click the **Next** command button.

7 OPERATION - FRESH START

17) Set Recording Parameters such as the instrument name, recording interval, start parameters and stop parameters. This step is where the user can also turn a sensor channel **ON** or **OFF**, set the sensor location description and sensor type.



The M.O.L.E.[®] Thermal Profiler should be programmed to record data every 0.1 seconds which is optimal for solar metallization processes.

Fresh Start

Instrument Name: M.O.L.E. Thermal Profiler

Recording Interval
Hour: 0 Minute: 0 Second: 0 1/10: 1

Start Parameters
☐ Temperature 33.0
Trigger Slope: Positive (+)
☐ Delay Points 1
Delay Time: 00:00:00

Stop Parameters
☒ Data Points 600
Total Time: 0000:00:01:00
☒ Synchronize instrument clock

Channel	On	Location	Type	Gain	Offset
1	☒	Sensor 1 Location	Type-K	1.00000	0.00000
2	☒	Sensor 2 Location	Type-K	1.00000	0.00000
3	☒	Sensor 3 Location	Type-K	1.00000	0.00000


Help << Previous Next >> Finish Cancel

18) Click the **Next** command button.

19) Verify the instrument status. This dialog box displays the health of the M.O.L.E. Thermal Profiler such as battery charge, internal temperature, thermocouple temperatures.



If everything is **OK**, the dialog box displays a **GREEN** sign. If there are any items that may prevent the user from collecting good data, they are highlighted and a **RED** sign is displayed.



20) Click the **Next** command button.


21) Review the oven settings and click the **Next** command button to continue.

Step 2: Perform Data Run

1) Place the M.O.L.E.[®] Thermal Profiler in the thermal barrier.



Never permit the M.O.L.E.[®] Thermal Profiler to exceed the absolute maximum warranted internal temperature, as permanent damage may result. The thermal barrier supplied with the system is specifically designed to protect the M.O.L.E.[®] through typical solar metallization processes.



2) After the oven stabilizes:

- Press the **ON/OFF** button.
- Press the **Record** button.
- Close the thermal barrier making sure it is fully latched and the thermocouple sensor wires are not damaged.



The **Record** button will need to be pressed even if the M.O.L.E.[®] Thermal Profiler is configured to begin recording data if start parameters **Trigger Temperature** or **Points Delay** are configured.

3) Pass the thermally protected M.O.L.E.[®] Thermal Profiler and test solar cell through your process.




It is highly recommended that protective gloves are used when retrieving the thermal barrier from the oven and when opening the thermal barrier.

4) As the test solar cell and M.O.L.E.[®] Thermal Profiler emerge from the machine carry to a table or flat surface.

5) Open the thermal barrier and if the **Record LED** is still flashing this means the M.O.L.E.[®] Thermal Profiler is still recording and it must be stopped by pressing the **Record** button.



If a thermocouple sensor is removed before the M.O.L.E.[®] Thermal Profiler has stopped recording data, the data for that channel will show as OPEN T/C.




6) Remove the M.O.L.E.[®] Thermal Profiler from the thermal barrier. Handle it carefully, as the case may still be warm.

7) Disconnect the thermocouple sensors from the M.O.L.E.[®] Thermal Profiler.

Step 3: Read Data Run


1) Restore the software and click the **Finish** command button to read the data run from the M.O.L.E.[®] Thermal Profiler.



On this step of the wizard, the user has the ability to remove a selected data run from the V-M.O.L.E.[®] Thermal Profiler by either selecting the **Delete After Reading** check box or selecting the **Delete** command button and removing it before downloading.


B OPERATION - FRESH START

2) When the data run has been downloaded, the software will prompt the user to save the data run file (*.XMG).



3) When finished, click the **Save** command button.

4) The software then prompts the user if they want to enter Prediction mode. Entering prediction mode enables the user to change a zone temperature values or the conveyor speed and predict the outcome of that change on the data run profile.



5) Click the **Predict** command button to enter Prediction mode or **Done** to complete the workflow wizard.

The information is automatically saved in the data run file (*.XMG) and the experiment data can now be analyzed with the software tools.



SINCE 1964

World Headquarters

ECD (N. America and Europe)
4287-B SE International Way
Milwaukie, Oregon 97222-8825
U.S.A

Telephone: +(1) 503.659.6100
 +(1) 800.323.4548

Fax: +(1) 503.659.4422

E-Mail: ecd@ecd.com

Web: ecd.com

Asia

ECD Asia/Pacific

Singapore

Mobile: +(65) 9692.6822

E-Mail: ecd.asia@ecd.com

© 2010 ECD. All Rights Reserved. Foreign and US Products of ECD are covered by US Patents and Patents Pending.

The trapezoidal ECD logo®, and M.O.L.E.® (Multi-Channel Occurrent Logger Evaluator) are registered trademarks of ECD.